

EU Declaration of Conformity
according to Article 19
Medical Device Regulation 2017/745/EU

KAVO

Manufacturer

KaVo Dental GmbH
Bismarckring 39
88400 Biberach
Germany
www.kavo.com

EUDAMED SRN

DE-MF-000006471

Product / REF

QUATTROcare PLUS 2124 A / 1.008.3805

Basic UDI-DI

++EKAVG515ZE

Classification

Class I, Rule 13

Intended use of the product (s)

The intended use of the KaVo QUATTROcare is for internal cleaning, i.e, purging of old lubricant, for the maintenance of rotating dental and surgical instruments. KaVo QUATTROcare spray can is a component of the KaVo QUATTROcare unit and is specifically designed to fit this unit only. The lubricant spray can is unable to be used manually.

For detailed description of product and accessories see instructions for use

EU Marking in accordance with

Regulation on medical devices (MDR)

2017/745/EU

Common Specifications

Currently not available

Statement

We declare under our sole responsibility that the products manufactured by us to which this declaration relates conform to the essential safety and performance requirements in accordance with the provisions of the regulation above and their applicable annexes.

This declaration is supported by the certificate with registration no. 51512-21-01 according to application of a QM system according to EN ISO 13485:2016 + AC:2018 + A11:2021.


Validity

Issued on

2025-04-01

Valid until

2027-09-28


Klaus Reisenauer
Senior Director Regulatory Affairs
Quality Assurance